

## UNITED STATES ARTMENT OF COMMERCE Patent and Trademark Office

Address: COMMISSIONER OF PATENTS AND TRADEMARKS Washington, D.C. 20231

APPLICATION NO.	FILING DATE	FIRST NAMED	INVENTOR	ATTO	ORNEY DOCKET NO.
08/753	,288 11/2:	2/96 BATTIATO		D	L-F/168
_		QM31/0	524 ¬	EXA	MINER
	W HUMPHREY ERRON & EVAI			FINKEL,S	
	AREW TOWER		ART	UNIT	PAPER NUMBER
CINCIN	NATI OH 452	<del>, , _</del>		3734	5
		. Р	DATE MA	NLED:	_

Please find below and/or attached an Office communication concerning this application or proceeding.

**Commissioner of Patents and Trademarks** 



## Office Action Summary

Application No. **08/753,288** 

Applicant(s)

Battiato et al.

Examiner

**Sharon Finkel** 

Group Art Unit 3734

Ľ	al.
	<u>. I. Janusuds 1200). Massak suntun 300 juuri 301 suud</u>
_	·

<ul> <li>□ This action is FINAL.</li> <li>□ Since this application is in condition for allowance except for formal matin accordance with the practice under Ex parte Quayle, 1935 C.D. 11; 4</li> <li>A shortened statutory period for response to this action is set to expire is longer, from the mailing date of this communication. Failure to respond application to become abandoned. (35 U.S.C. § 133). Extensions of time 37 CFR 1.136(a).</li> <li>Disposition of Claims</li> </ul>	153 O.G. 213.  O month(s), or thirty days, whichever within the period for response will cause the may be obtained under the provisions of  is/are pending in the application.
in accordance with the practice under <i>Ex parte Quayle</i> , 1935 C.D. 11; 4 A shortened statutory period for response to this action is set to expire is longer, from the mailing date of this communication. Failure to respond application to become abandoned. (35 U.S.C. § 133). Extensions of time 37 CFR 1.136(a).  Disposition of Claims	153 O.G. 213.  O month(s), or thirty days, whichever within the period for response will cause the may be obtained under the provisions of  is/are pending in the application.
is longer, from the mailing date of this communication. Failure to respond application to become abandoned. (35 U.S.C. § 133). Extensions of time 37 CFR 1.136(a).  Disposition of Claims	within the period for response will cause the may be obtained under the provisions of  is/are pending in the application.
[77] OLI 100 4 00	
	in form with drawn from a posidoration
Of the above, claim(s)	is/are withorawn from consideration.
☐ Claim(s)	is/are allowed.
☐ Claim(s)	is/are rejected.
Claim(s)	is/are objected to.
	ubject to restriction or election requirement.
Application Papers  See the attached Notice of Draftsperson's Patent Drawing Review, P The drawing(s) filed on is/are objected to by th The proposed drawing correction, filed on is The specification is objected to by the Examiner. The oath or declaration is objected to by the Examiner.  Priority under 35 U.S.C. § 119 Acknowledgement is made of a claim for foreign priority under 35 U All Some* None of the CERTIFIED copies of the priority received. Teceived in Application No. (Series Code/Serial Number) Teceived in this national stage application from the International *Certified copies not received: Acknowledgement is made of a claim for domestic priority under 35	e Examiner.  _approved _disapproved.  .S.C. § 119(a)-(d). y documents have been  al Bureau (PCT Rule 17.2(a)).
Attachment(s)  Notice of References Cited, PTO-892 Information Disclosure Statement(s), PTO-1449, Paper No(s). Interview Summary, PTO-413 Notice of Draftsperson's Patent Drawing Review, PTO-948 Notice of Informal Patent Application, PTO-152	

Art Unit: 3734

## Restriction

1. Restriction to one of the following inventions is required under 35 U.S.C. 121:

2. This application contains claims directed to the following patentably distinct species of the claimed invention:

Species Group 1, claims 1-7, directed to an injector having a control circuit responsive to air detection means;

Species Group 2, claims 11-20, directed to an injector having a control circuit responsive to hand-operated movement control;

Species Group 3, claims 21-26, 28-30 and 32, directed to an injector having a control circuit responsive to an electronic display of information which is in turn responsive to a tilt angle sensor;

Species Group 4, claims 27 and 31, directed to an injector having a control circuit responsive to an electronic display of information which is in turn responsive to a tilt angle sensor and the control circuit is also responsive to hand-operated movement control;

Species Group 5, claim 33, directed to an injector having a control circuit responsive to a face plate mounting and face plate mounted thereon, and the magnetic fields generated thereby;

Species Group 6, claims 34-38, directed to an injector having a control circuit responsive to hand-operated movement control and an encoder controlling the motor on said plunger drive.

Art Unit: 3734

3. Applicant is required under 35 U.S.C. 121 to elect a single disclosed species for prosecution on the merits to which the claims shall be restricted if no generic claim is finally held to be allowable. Currently, no claim is generic. For clarification, the following elements are deemed to be generic to all claims listed above: plunger drive ram, motor for moving said plunger drive ram, syringe, and control circuit.

- 4. Applicant is advised that a response to this requirement must include an identification of the species that is elected consonant with this requirement, and a listing of all claims readable thereon, including any claims subsequently added. An argument that a claim is allowable or that all claims are generic is considered nonresponsive unless accompanied by an election.
- 5. Upon the allowance of a generic claim, applicant will be entitled to consideration of claims to additional species which are written in dependent form or otherwise include all the limitations of an allowed generic claim as provided by 37 CFR 1.141. If claims are added after the election, applicant must indicate which are readable upon the elected species. MPEP § 809.02(a).
- 6. Should applicant traverse on the ground that the species are not patentably distinct, applicant should submit evidence or identify such evidence now of record showing the species to be obvious variants or clearly admit on the record that this is the case. In either instance, if the examiner finds one of the inventions unpatentable over the prior art, the evidence or admission may be used in a rejection under 35 U.S.C. 103(a) of the other invention.

Art Unit: 3734

7. Applicant is advised that the reply to this requirement to be complete must include an election of the invention to be examined even though the requirement be traversed (37 CFR 1.143).

- 8. Applicant is reminded that upon the cancellation of claims to a non-elected invention, the inventorship must be amended in compliance with 37 CFR 1.48(b) if one or more of the currently named inventors is no longer an inventor of at least one claim remaining in the application. Any amendment of inventorship must be accompanied by a petition under 37 CFR 1.48(b) and by the fee required under 37 CFR 1.17(I).
- 9. The following additional restriction is required under 35 U.S.C. 121
  - Group I. Claims 1-7, drawn to an injector with an air detector for detecting the presence of air contained in a syringe, a warning signal and a circuit responding to the warning, classified in class 604, subclass 122.
  - Group II. Claims 8-10, drawn to a syringe having an means (lens) for detecting air in the syringe, classified in class 604, subclass 122.
- 10. The inventions are distinct, each from the other because of the following reasons:

  Inventions I and II are related as combination and subcombination. Inventions in this relationship are distinct if it can be shown that (1) the combination as claimed does not require the particulars of the subcombination as claimed for patentability, and (2) that the subcombination has utility by itself or in other combinations (MPEP § 806.05(c)). In the instant case, the combination as claimed does not require the particulars of the subcombination as claimed because the air detector

Art Unit: 3734

as claimed can detect air anywhere "contained within said syringe", and the syringe has an "outwardly-projecting section positioned on the discharge tip." The subcombination has separate utility such as providing means for the detection of air or medicine entering the needle by manual methods (by eye), to assist in the proper filling of the syringe.

11. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Sharon Finkel whose telephone number is (703) 305-0154.

sef

June 15, 1998

Stinkel

RONALD STRIGHT